

LIQUIZYMÉ

TOTAL BILIRUBIN

(DMSO Method)

**BEACON**

Code	Product Name	Pack Size
LS054A	Liquizyme Total Bilirubin	240 ml
LS054B	Liquizyme Total Bilirubin	1200 ml

Intended Use

Diagnostic reagent for quantitative *in vitro* determination of Bilirubin in human serum.

Clinical Significance

Bilirubin is a breakdown product of haemoglobin. Bilirubin formed in the reticulo endothelial system is transported bound by albumin to the liver. This bilirubin is water insoluble and is known as indirect or unconjugated bilirubin. In the liver, bilirubin is conjugated to glucuronic acid to form direct bilirubin. Conjugated bilirubin is excreted via the biliary system into the intestine. Here it is metabolised by bacteria to urobilinogen & stercobilinogen.

Principle

In the determination of Bilirubin Total, Bilirubin is coupled with diazotized sulphanilic acid in the presence of ethylene glycol and dimethylsulfoxide as solvents to produce an intensely colored diazo dye. The intensity of colour of this solution is proportional to the concentration of the bilirubin total in the sample.

Reaction**Total Bilirubin**

Bilirubin + Sulphanilic acid + Sodium Nitrite $\xrightarrow{\text{DMSO}}$ Azobilirubin

Contents:**Reagent 1 : Total Bilirubin Reagent**

Buffer : < 15 mmol/l
Sulphanilic Acid : >20 mmol/l

Reagent 2 : Total Nitrite Reagent

Sodium Nitrite : > 10 mmol/l

Reagent 3: Bilirubin Artificial Standard : 10 mg/dl

Ready to use

Unit Conversion

mg/dl x 16.95 = μ mol/l

Normal Value :

Serum Total Bilirubin : upto 1.0 mg/dl

Each Laboratory should establish it's own normal range representing its patient population.

Reagent Preparation

Reagents are liquid, ready to use.

Stability and storage

The unopened reagents are stable till the expiry date stated on the bottle and label when stored at room temperature.

Total Bilirubin**Performance Data**

Data contained within this section is representative of performance on Beacon system. Data obtained in your laboratory may differ from these values.

Limit of quantification : 0.0052 mg
Linearity : 20 mg/dl
Measuring range : 0.0052 – 20 mg/dl

Intra-assay precision Within run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	1.06	0.03	3.16
Sample 2	4.47	0.04	0.92

Inter-assay precision Run to run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	4.06	0.01	0.35

Comparison

A comparison between Beacon Total Bilirubin (y) and a commercially available test (x) using 20 samples gave following results:

y = 0.999 x + 0.042 mg/dl
r = 0.999

Linearity:

This procedure is linear upto 20 mg/dl. If the values exceed this limit, dilute the sample with normal saline (NaCl 0.9%) and repeat the assay. Multiply result by dilution factor.

Interferences

Following substances do not interfere:
haemoglobin up to 7.5 g/l, triglycerides up to 1500 mg/dl.

Warning And Precautions

For *in vitro* diagnostic use. To be handled by entitled and professionally educated person.

Procedure:

Pipette into clean dry test tubes labeled as **Blank (B)**, and **Test (T)** :

Total Bilirubin

Addition Sequence	(B)	(T)
Total Bilirubin Reagent	1000 µl	1000 µl
Total Bilirubin Activator	-	20 µl
Serum	50 µl	50 µl

Mix well and incubate at 37°C for exactly 5 minutes. Measure the absorbance of the Test Samples (Abs. T) immediately against their respective Blanks.

Calculation**With factor:**

Total Bilirubin = OD of test - OD of sample blank x Factor (20)

With artificial standard:

$$\text{Bilirubin Concentration (mg/dl)} = \frac{\text{OD test} - \text{OD of sample Blank}}{\text{OD of standard}} \times 10$$

Assay Parameters For Photometers

Mode	End Point
Wavelength	546
Sample Volume (µl)	50
Total Bilirubin Reagent (µl)	1000
Total Bilirubin Activator (µl)	20
Incubation Time	5 min
Reaction temperature (°C)	37
Linearity Low (mg/dl)	0.0052
Linearity High (mg/dl)	20
Blank with	Serum
Standard Concentration	10 mg/dl
Unit	mg/dl

References

1. Cornall, A. G., Bardawill, C. J., David, M. M.: J. Biol. Chem. 177, 751, 1949.
2. Doumas, B. T., Bayse, D. D. a kol.: Clin. Chem. 27, 1642, 1981.
3. Chromý, V., Fischer, J.: Clin. Chem. 23, 754, 1977.
4. Chromý, V., Fischer, J., Vozníček, J.: Z. Med. Labor. Diagn. 21, 333, 1980.
5. Tietz Textbook of Clinical Chemistry and Molecular diagnostics. Burtis, C.A.
6. Ashwood, E.R., Bruns, D.E.; 5th edition, WB Saunders.

Symbols Used On Labels

Catalogue
Number



Manufacturer



See Instruction
for Use



Lot Number



Content



Storage Temperature



Expiry Date



In Vitro Diagnostics

BEA/24/BTD/LS/IFU Ver-01
23/07/2024

