

HDL DIRECT SYSTEM PACK

(PEGME METHOD)

B Auto 200, Unicorn 230, Unicorn 120 , Bonavera Chem 200 ,
Beaconnic chem 200, Beaconnic B200, Beaconnic analyzer 120 &
Bonavera chem 100(Fully Auto Biochemistry Analyzer)



Code	Product Name	Pack Size
BA221	HDL Direct System Pack	4x30 + 4x10 ml
BA221A	HDL Direct System Pack	2x30 + 2x10 ml

INTENDED USE

Diagnostic reagent for quantitative *in vitro* determination of HDL Cholesterol in human serum and plasma.

CLINICAL SIGNIFICANCE

High-density lipoproteins (HDL) compose one of the major classes of plasma lipoproteins. They are synthesized in liver as complexes of apolipoprotein and phospholipid and are capable of picking up cholesterol and carrying it from arteries to the liver, where the cholesterol is converted to bile acids and excreted into the intestine.



An inverse relationship between HDL-cholesterol (HDL-C) levels in serum and the incidence/prevalence of coronary heart disease (CHD) has been demonstrated in a number of epidemiological studies. The importance of HDL-C as a risk factor for CHD is now recognized.

Accurate measurement of HDL-C is of vital importance when assessing patient's risk for CHD.

PRINCIPLE

The assay is based on a modified polyvinyl sulfonic acid (PVS) and polyethyleneglycol-methyl ether (PEGME) coupled classic precipitation method with the improvements in using optimized quantities of PVS/PEGME and selected detergents. LDL, VLDL and chylomicron (CM) react with PVS and PEGME and the reaction results in inaccessibility of LDL, VLDL and CM by cholesterol oxidase (CHOD) and cholesterol esterase (CHER).

The enzymes selectively react with HDL to produce H₂O₂ which is detected through a Trinder reaction.



REAGENT COMPOSITION

Reagent 1: R1 Reagent

Buffer	> 5 mmol/l
MgCl ₂	> 2 mmol/l
TOOS	< 2 mmol/l

Reagent 2 : R2 Reagent

CHE	> 2 U/L
COD	< 5 KU/L
POD	< 10 KU/L

Reagent 3: Ultima HDL Calibrator

Concentration see on label

REAGENT PREPARATION

Reagents R1 and R2 are liquid, ready to use. Calibrator reconstitute with 1 ml of deionized water at +20-+25°C and mix gently (avoid foaming). Allow to stand for at least 30 minutes until complete reconstitution before use. Store reconstituted calibrator at +2-+8°C.

STABILITY AND STORAGE

The unopened reagents are stable till the expiry date stated on the bottle and kit label when stored at +2-+8°C.

Reagents are light-sensitive. Do not let bottles remain open. Keep containers tightly closed.

The reconstituted calibrator is stable for 7 days at +2 - +8°C.

On board stability : Min 30 days If refrigerated (+8 - +14°C) and not contaminated.

SPECIMEN COLLECTION AND HANDLING

Use serum or heparin plasma.

It is recommended to follow NCCLS procedures (or similar standardized conditions)

Stability in serum/plasma:	24 hours	at +20 - +25°C
	7 days	at +4 - +8°C

Discard contaminated specimens.

CALIBRATION

Calibration with Ultima HDL calibrator is recommended.

QUALITY CONTROL

It's recommended to run normal and abnormal control sera to validate reagent performance.

NORMAL VALUE

Adult Male	35 - 80 mg/dl
Adult Female	42 - 88 mg/dl

It is recommended that each laboratory verify this range or derives reference interval for the population it serves.

PERFORMANCE DATA

Data contained within this section is representative of performance on Beacon. Data obtained in your laboratory may differ from these values.

Limit of quantification: 2.32 mg/dl

Linearity: 180 mg/dl

Measuring range: 2.32 - 180 mg/dl

Intra-assay precision Within run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	61.10	1.45	2.37
Sample 2	97.20	1.11	1.14
Inter-assay precision Run to run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	35.73	0.70	1.97

COMPARISON

A comparison between HDL System Pack (y) and commercially available test (x) using 20 samples gave following results:

$$y = 0.991x + 0.851 \text{ mg/dl}$$

$$r = 0.999$$

INTERFERENCES

Following substances do not interfere:

haemoglobin up to 10 g/l, bilirubin up to 40 mg/dl, triglycerides up to 2000 mg/dl Interference by N-acetylcysteine (NAC), acetaminophen and metimazole causes falsely low results. To carry out the test, blood withdrawal should be performed prior to administration of drugs.

WARNING AND PRECAUTIONS

For *in vitro* diagnostic use. To be handled by entitled and professionally educated person. Reagent of the kit are not classified like dangerous.

Serum used in the manufacture of the calibrator has been tested by FDA-approved methods and found non-reactive for hepatitis B surface antigen (HbsAg), antibody to Hepatitis C (HCV), HIV-1 p24 antigen and antibody to HIV 1/2. The test procedures do not guarantee that all infectious agents will be detected. Because no test method can offer complete assurance that Hepatitis B virus Hepatitis C virus and HIV 1/2 or other infectious agents are absent, the material should be handled as potentially infectious. MSDS will be provided on request.

WASTE MANAGEMENT

Please refer to local legal requirements.

Parameter For B Auto 200, Unicorn 230, Unicorn 120 & Bonavera Chem200, Beaconic chem 200, Beaconic B200, Beaconic analyzer 120, Bonavera chem 100 (Fully Auto Biochemistry Analyzer)

Test Name	HDL DIRECT
Full Name	HDL DIRECT
PRI Wave	578 nm
SEC Wave	700 nm
Assay/Point	1 POINT END
Start	-
End	34
Decimal	2
Unit	mg/dl
Linearity Range Low	2.32
Linearity Range High	180
Sample Volume	2 µl

Reagent 1 (R1) Volume	150 µl
Reagent 1 (R2) Volume	50 µl
Substrate Depleted/Abs.limit	-
Linearity	180 mg/dl
Out Of Linearity Range	-
Calibration Type	2 Point linear
Points	2
Blank Type	Reagent
Concentration Blank	0.00
Concentration Std	Refer calibrator label

NOTE

The program is made as per the in house testing, it can be modified as per requirements. Clinical diagnosis should not be made on findings of a single test results, but both clinical and laboratory data.

REFERENCES

- Dominiczak M, McNamara J. The system of Cardiovascular prevention. 103- 125; Nauk M, Wiebe D. Warnick G. Measurement of High-Density-Lipoprotein Cholesterol. 221-224. In: Handbook of Lipoprotein Testing (eds. Rifai, Warnick and Dominiczak) 2nd edition.
- Bar, D.P., Russ E.M, Eder, H.A. Protein-lipid relationships in human plasma, Am. J. Med., 11;480(1951).
- Gordon, T. et al., High density lipoprotein as a protective factor against coronary heart disease, Am. J. Med. 62,707 (1977).
- Castelli, W. P et al., HDL Cholesterol and other lipids in coronary heart disease, Circulation, 55,767 (1977).
- National Institutes of Health publication No. 93-3095, September (1993).
- Williams, P., et al., High density lipoprotein an add coronary risk factor, Lancet, 1,72(1979).
- Kannel, w.B., Castelli W.P., Gordon, T. Cholesterol in the prediction of atherosclerotic disease; new perspectives based on the Framingham study, Ann. Intern. Med., 90:85, (1979).
- Castelli, W.P., et al, Cholesterol and other lipids in coronary heart disease.
- Pisani T, GebSKI CP, Leary Et, et al. Accurate Direct Determination of Low- Density Lipoprotein Cholesterol Assay. Arch Pat Lab Med 1995; 119:1127.
- Tietz textbook of Clinical Chemistry and Molecular Diagnostics. Burtis, C. A, Ashwood, E.R., Bruns, D.E; 5th edition, WB Saunders Comp., 2012.

Symbols Used On Labels



Catalogue
Number



Manufacturer



See Instruction
for Use



Lot Number



Content



Storage Temperature



Expiry Date



In Vitro Diagnostics

BEA/24/HDL/SB/IFU ver-05
28/06/2025

