

FEBRILE ANTIGEN

(WIDAL TUBE METHOD)

Code	Product Name	Pack Size
SE005A	Febrile Antigen(Widal Tube Method)	4 x 50 ml
SE005B	Febrile Antigen(Widal Tube Method)	2 +2 x 50 ml

INTENDED USE :

This diagnostic reagent kit is used for detection of specific antibodies present in response to the stimulation by specific antigen of Salmonella (group).

PRINCIPLE :

The killed bacterial suspension of Salmonella carries specific 'O' and 'H' antigen will react with immuno specific antibodies which may be present in patient serums and agglutinate the antigen, and produce agglutinin or clumps in the tube.

CLINICAL SIGNIFICANCE :

The organism Salmonella typhosa is responsible for producing the disease in mankind which is known as enteric fever or typhoid fever, which is characterized generally by very high constant fever, loss of appetite, transitory bacteraemia, round or oval shaped ulcer with smooth peritoneal surface of Peyer's patches and solitary lymphoid follicles of ileum etc. The organism possess 'O' antigen on the cell wall and 'H' antigen on its flagella, against which the host body produces Immunospecific antibodies, to counteract the effect of corresponding antigens. On the other hand the paratyphoid fever caused by Salmonella paratyphi A or paratyphi B is characterized by milder course of disease. These organisms also possess somatic 'O' and flagellar antigen which is termed as 'A(H)' and 'B(H)' respectively. The other organisms of Salmonella species like S. typhimurium responsible for causing food poisoning or Arisona group causing fetal infection do have similar antigenic properties.

Contents :

Reagent 1 : Febrile S.typhi "O" Antigen
Reagent 2 : Febrile S.typhi "H" Antigen
Reagent 3 : Febrile S.paratyphi "A(H)" Antigen
Reagent 4 : Febrile S.paratyphi "B(H)" Antigen

SAMPLE :

Fresh serum sample is preferred. In case of any delay the sample should be stored at 2°- 8°C away from direct light. However the test is to be performed within 24 hrs. of collection of sample.

STORAGE AND STABILITY :

All reagents are stable till expiry date mentioned on the label when stored at 2°-8°C away from direct light.

PROCEDURE :

- For each serum sample under test, arrange four rows of 6 tubes each (3" x 3/8") in a rack.
- Prepare master dilution by taking 5 tubes (5" x 5/8") in another rack. Pipette 7ml of normal saline (0.85% Sodium chloride) in the first tube and 3.5ml in each of the remaining four tubes. Add 0.5ml of serum to the first tube and mix well. Continue successive transfer of 3.5ml from quantities till the last tube is reached. This will give final dilution of 1:30, 1:60, 1:120, 1:240 and 1:480.
- Transfer 0.5ml quantities from the master dilution tubes to each tube of the corresponding vertical row in test rack. Place 0.5ml of normal saline to each of the tubes in the Last (i.e. 6th) row to serve as controls.
- To each of the six tube in the first, second, third and fourth horizontal row, add 0.5ml of S. typhi 'O', S. typhi 'H', S.paratyphi A (H) and S. Paratyphi B (H) antigens respectively.
- Shake the rack well to mix and incubate at 37°C for overnight (16-20 hours).

- Note the highest dilution in which there is evidence of agglutination as observed by naked eye or a hand lens. With 'H' antigen the pattern of agglutination is floccular cotton wool type whereas with 'O' antigen it is fine granular and mat type at the bottom of the tube.
In addition to the pattern of sedimented organisms, the decrease in opacity of the supernatant as compared to the Saline control tube must be observed and taken into account while judging the degree of agglutination.

INTERPRETATION OF RESULT :

- Sera from normal individuals may agglutinate these antigens, level of "normal" agglutinin varies in different countries and different communities.
- Agglutination titres of 1:120 and more are significant and rise in the titres on repetition of the test after a few days will confirm the diagnosis of enteric fever.

LIMITATIONS

Quantitative tube test is non specific type of test. In case of doubt, the results should be further confirmed by microbiological Tests, microscopy, histopathology etc.

TO REMEMBER

- Use clean glassware free from dust or debris.
- Care should be taken to maintain the temp. of 37°C through out the period of Incubation.
- Agglutination titres of 1:240 and above are typically found in cases of enteric fever. The Specific organism responsible is Determined by noting the 'H' agglutinin titre.
- Anamnestic Reaction: Persons who have suffered from enteric infection in past or who had received TAB vaccine may show appearance of agglutinins in moderate titre when suffering from other unrelated illness. Such anamnestic appearance of agglutinins can be differentiated from true infection by demonstrating the marked rise in the titre when the test is repeated after a few days.
- TAB Vaccination : A moderated rise in titre of all three 'H' agglutinins simultaneously against all antigens is suggestive of recent TAB Vaccination. A careful history-taking will easily clarify the point.

REFERENCES :

- Felix A. (1942). Brit Med. Jr. 11,597.
- Protell R.I.E.et.al. (1971) Lancet, 11,330.
- Medical Bacteriology. N. C. Dey (1970) 259-284



REF	Catalogue Number	Manufacturer	See Instruction for Use
LOT	Lot Number	CONT Content	Storage Temperature
Expiry Date	IVD In Vitro Diagnostics		

BEA/24/FEB/SE/IFU-00